

Research MRI Safety Committee Standard

REPORTING REQUIREMENTS

Category: Magnetic Resonance Imaging (MRI) Safety

Procedure #: MR.SOP.10

Applies to: Investigators, study personnel, Medical College of Wisconsin (MCW) staff

PURPOSE:

The mission of the Research MRI Safety Committee at the Medical College of Wisconsin (MCW) Office of Research and Clinical Programs is to protect the health and safety of research subjects and staff involved in research procedures using MRI technology. (c)2018 (e)2.5 (t)2.5 (es)2.5

Research MRI Safety Committee and administration

2. Timeframe of the report:
 - a. Within 1 hour of occurrence, if the damage/failure affects the primary function of the MRI System
 - b. Within 8 hours of occurrence, if the damage/failure affects the ancillary function of the MRI System
 - c. Within 24 hours of occurrence, if the damage/failure affects components of the MRI System that do not affect the system's ability to image
3. Must be reported by: Study Personnel if discovered during a study

Approval Date: 02/15/2024
Effective Date: 02/15/2024
Review Date: 02/15/2024
Revision History: 01/2009, 12/17/2012, 11/13/2014